

**Notice of Allowability**

Application No.

09/913,617

Applicant(s)

HERRMANN, INGO F.

Examiner

John P. Leubecker

Art Unit

3739

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to March 25, 2004.
2. ☒ The allowed claim(s) is/are 25,26,29,30,34-48 and 51-53.
3. ☒ The drawings filed on 1/8/02 are accepted by the Examiner.
4. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All    b) ☐ Some\*    c) ☐ None    of the:
    1. ☒ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  6. ☐ CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).**
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

- |   |  |
|---|--|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892)  | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)            |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                | 6. <input type="checkbox"/> Interview Summary (PTO-413),<br>Paper No./Mail Date _____. |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),<br>Paper No./Mail Date _____ | 7. <input type="checkbox"/> Examiner's Amendment/Comment                               |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit<br>of Biological Material          | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance   |
|   | 9. <input type="checkbox"/> Other _____.   |

John P. Leubecker  
Primary Examiner  
Art Unit: 3739

### **REASONS FOR ALLOWANCE**

1. The following is an examiner's statement of reasons for allowance: Claim 25, as amended in the response filed March 25, 2004, now recite a combination of elements that reads over the prior art of record. Although Adair ('175) allows for longitudinal movement of the instrument with respect to the fiberscope part, the device is designed such that one magnet fixedly holds the other magnet, wherein the part of the holding device that accommodates the instrument is not explicitly disclosed as being capable of longitudinal displacement with respect to the part of the holding device on the fiberscope part when these parts are joined as a unit. In other words, the magnets of Adair are not disclosed as capable of longitudinally moving relative to one another when magnetically attached. Since no other references of record explicitly teach or fairly suggest combination as now defined in claim 25, claim 25 is allowable over the prior art of record.


Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Leubecker whose telephone number is (703) 308-0951. The examiner can normally be reached on Monday through Friday, 6:00 AM to 2:30 PM.

Art Unit: 3739

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C.M. Dvorak can be reached on (703) 308-0994. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



John P. Leubecker  
Primary Examiner  
Art Unit 3739

jpl

### **REASONS FOR ALLOWANCE**

1. The following is an examiner's statement of reasons for allowance: Claim 25, as amended in the response filed March 25, 2004, now recite a combination of elements that reads over the prior art of record. Although Adair ('175) allows for longitudinal movement of the instrument with respect to the fiberscope part, the device is designed such that one magnet fixedly holds the other magnet, wherein the part of the holding device that accommodates the instrument is not explicitly disclosed as being capable of longitudinal displacement with respect to the part of the holding device on the fiberscope part when these parts are joined as a unit. In other words, the magnets of Adair are not disclosed as capable of longitudinally moving relative to one another when magnetically attached. Since no other references of record explicitly teach or fairly suggest combination as now defined in claim 25, claim 25 is allowable over the prior art of record.


Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Leubecker whose telephone number is (703) 308-0951. The examiner can normally be reached on Monday through Friday, 6:00 AM to 2:30 PM.

Art Unit: 3739

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C.M. Dvorak can be reached on (703) 308-0994. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



John P. Leubecker  
Primary Examiner  
Art Unit 3739

jpl